



ANNUAL X-RAY INSPECTION REPORT 2013 RADIOLOGICAL HEALTH

108 Cherry Street, PO Box 70, Burlington, VT 05402-0070

TABLE OF CONTENTS

| | | |
|------------------------------------|-------|----|
| Executive Summary | ----- | 2 |
| Overview | ----- | 3 |
| Summary of All Inspections | ----- | 10 |
| Dental Inspections | ----- | 13 |
| Non-compliance Items | | |
| Dose to Patient per Exposure | | |
| Annual Dose to Occupational Worker | | |
| Annual Dose to Public | | |
| Medical Inspections | ----- | 17 |
| Non-compliance Items | | |
| Dose to Patient per Exposure | | |
| Annual Dose to Occupational Worker | | |
| Annual Dose to Public | | |
| Chiropractic Inspections | ----- | 21 |
| Non-compliance Items | | |
| Dose to Patient per Exposure | | |
| Annual Dose to Occupational Worker | | |
| Annual Dose to Public | | |
| Podiatric Inspections | ----- | 24 |
| Non-compliance Items | | |
| Dose to Patient per Exposure | | |
| Annual Dose to Occupational Worker | | |
| Annual Dose to Public | | |
| Veterinary Inspections | ----- | 27 |
| Non-compliance Items | | |
| Exposure to Patient per Exam | | |
| Annual Dose to Occupational Worker | | |
| Annual Dose to Public | | |

EXECUTIVE SUMMARY

A total of 127 x-ray facilities were inspected in 2013. Out of the 127 facilities 71 (56%) were in full compliance at the time of the inspection. Forty five (80%) of those facilities not in compliance came into compliance after the inspection.

The National Council on Radiation Protection and Measurements (NCRP) recommends that medical facilities (which includes chiropractic facilities) be inspected every two years. Dental and veterinary facilities are recommended to be inspected every 4 years. Podiatric facilities are inspected every 4 years since the x-ray machines are similar to dental x-ray machines.

The main area of concern is the lack of satisfactory lead aprons. Lead aprons are available in the facilities but some are cracked or torn. The facilities are encouraged to obtain new lead aprons and check them annually. Other non-compliance items are listed on pages 5 to 9.

Annual dose rates to all operators of x-ray equipment of the facilities inspected were less than the maximum allowed limit of 5000 millirem, typically less than 1% of this limit. Annual dose rates to the public were less than the maximum allowed limits of 100 millirem, except for one facility. The dose to the public was measured at a distance of less than usual 6 feet due to the layout of the facility. The facility does not allow members of the public in that area during an x-ray examination.

The entrance skin exposure to the patient was within the appropriate limit for all facilities.

The dose to the patient and the dose to the operator is less for all x-ray facilities that use faster speed film. This can be observed most clearly for the dental facilities. As the speed of the film increases from "D" to "F" the average dose per exposure decreases from 0.47 to 0.26 millirem. It should also be noted that the use of digital x-ray again decreases the average dose per exposure from 0.26 millirem for "F" speed film to 0.15 millirem for direct digital x-rays. Doses from computed radiography (CR) are similar to doses from F speed film (0.25 and 0.26 millirem, respectively).

It is expected that as more digital x-rays are used we will see decreases in the total facility noncompliances because darkrooms, safelights, film, and processing are no longer needed. Fifty seven percent of dental, 44% of veterinary, 71% of medical, 7% of podiatric and 15% of chiropractic facilities are using digital x-ray. Fifty percent of all facilities are now using digital x-rays.

OVERVIEW

To be conservative, exposures to the operator and public are measured at the configuration of highest exposure possible. Exposure to the public is performed by aiming the x-ray tube out of the exam room door from approximately the patient position for an x-ray exam and measuring the exposure at the doorway where the public passes by in the hall. Operator exposures are measured at the position the operator stands when making the exposure as indicated by the facility.


Operator and public exposures are measured in milliroentgen per hour using a Victoreen 471 ion chamber. The exposure per hour is converted to annual dose by converting hours to year and milliroentgen to rem using the number of x-rays the facility takes within a given period of time. 1 milliroentgen is equal 0.5 millirem (American National Standard Institute 6.1.1-1991) for whole body exposure from scattered radiation for the public and operators.

Patient exposures are measured in milliroentgen using an Unfors Xi. Patient exposures are converted from milliroentgen to millirem using the following factors based on the organ of greatest risk. Multiplication of the factor by the number of milliroentgen per exam results in the dose in millirem.

| EXAM TYPE | FACTOR | ORGAN |
|-------------------|--------|---------------|
| Dental | 0.0015 | brain |
| PA Chest | 0.1044 | lung |
| AP Cervical Spine | 0.0435 | thyroid |
| AP Thoracic Spine | 0.1044 | lung |
| AP Lumbar Spine | 0.1044 | stomach/colon |
| AP Abdomen | 0.1044 | stomach/colon |
| AP Retrograde | 0.1044 | stomach/colon |
| Lateral Skull | 0.0218 | brain |
| Hand | 0.0087 | skin |
| Wrist | 0.0087 | skin |
| Arm | 0.1044 | bone marrow |
| Shoulder | 0.1044 | bone marrow |
| Leg | 0.1044 | bone marrow |
| Knee | 0.1044 | bone marrow |
| Ankle | 0.0087 | skin |
| DP Foot | 0.0087 | skin |
| Lateral Foot | 0.0087 | skin |

Adapted from National Council on Radiation Protection and Measurements Report No. 116 tissue weighting factors and conversion factor from roentgen to rad of 0.87 rad/roentgen.

The average radiation dose from natural and man-made sources is 620 millirem per year.



Adapted from NCRP Report No. 160, 2009, Ionizing Radiation Exposure of the Population of the United States.

INSPECTION ITEMS

The following boxed sections indicate the individual items that are specifically looked at during an inspection for the following general groups: film/screen, processing, darkroom/safelight, personnel monitoring, patient shielding, collimation, timer, kVp and filtration, patient entrance skin exposure criteria, public exposure criteria, operator conditions, and physical condition (x-ray unit, shielding, etc.)

Some inspection items may pertain only to specific types of facilities. For example, repeat rate analyses pertain only to chiropractic facilities, whereas panoramic units pertain only to dental facilities. There are also inspection items that cover all facilities (e.g., registration of all x-ray units).

New facilities are not cited for non-compliant items. However, they are given a period of approximately one month to correct any non-compliant items found in the initial inspection.

| | |
|--------------------|--|
| Film/screen | Dental film is less than E speed X-ray film speed is less than 400 Film is not protected from scatter radiation Film is not stored properly Film is exposed to chemicals Out of date film is used Film and screen types not matched No screen installation date is on outside of cassette Screen and cassettes are not of the same type or age Screen cleaning interval is inadequate Screen cleaning solution and lint free wipes are not used per manufacturer instructions Cassette check is inadequate Cassettes are not permanently identified for their type of use Film viewbox is not available Film viewbox is not cleaned periodically Viewbox bulbs are not of the same intensity and color Luminance of viewboxes is not similar Viewbox bulbs are not replaced annually Technique factors are not recorded in the patient log book Technique charts are not available or up to date Left/right markers are not used on clinical radiographs Clinical radiographs are not properly identified |
|--------------------|--|

| | |
|------------------------------------|--|
| Processing | <p>Thermometer is not available for manual processing Timer is not available for manual processing Floating cover is not present for manual processing Sight devevelopment is used No evidence of daily log is kept Developing technique recommended by the manufacturer is not used Developer and fixer temperature are not maintained in limits Processor cleaning interval is inadequate Processor is not operating properly Processor cleaning date is not recorded Clean-up film for processing all x-ray films (except intra-oral) are not run</p> |
| Darkroom/Safelight | <p>Safelight bulb is greater than 15 W Safelight is too close to the work area Light leaks are detected in the safelight housing Light leaks are detected in the safelight lens Safelight is improperly filtered Darkroom is not light tight Darkroom is not free of dust and dirt Daylight processor arm cuffs are not acceptable Daylight processor is not light tight Darkroom temperature/humidity are not acceptable There are other light sources present in the dark room</p> |
| Personnel Monitoring | <p>Personnel monitoring devices are required Control dosimeters are not properly used or stored Employee dosimeters are not properly used Employee dosimeters are not properly stored No evidence of employee review of records Personnel monitoring records are incomplete No radiation safety officer is designated for large practices Evidence of personnel holding film during exposure</p> |
| Personnel/Patient Shielding | <p>Satisfactory lead aprons are unavailable Satisfactory thyroid shields are unavailable Satisfactory gonadal shields are unavailable Lead aprons are improperly stored Lead aprons are not checked annually for tears and holes (radiographically or visually) Individuals holding patients are not protected Mobile equipment exposure switch cord is less than 6 feet long Non-essential individuals are in the x-ray room during exposure</p> |

Collimation
 X-ray beam is not restricted to the appropriate area
 X-ray beam is not restricted to the appropriate size
 Collimator light is not aligned with the x-ray field
 Collimation is not used in taking radiographs
 Collimator light is not bright enough under normal room lighting
 Collimator light problems (e.g. mirror broken, mirror obstructed)
 Inadequate collimation is used for clinical radiographs

Timer
 Timer does not terminate exposure
 Timer activates at zero
 Timer is inaccurate
 Timer repeatability is unacceptable
 No deadman switch is available

kVp and Filtration
 kVp is greater than 10% of set value
 kVp is non-repeatable
 Filtration in beam is less than required

Patient entrance skin exposure criteria (ESEC)
 ESEC in milliroentgen for non-specialty radiographic examinations shall not be exceeded when technical factors for an average adult patient are utilized:

| Examination | ESEC mR maximum | ESEC mR recommended | Body part thickness (cm) |
|---------------------------------|-----------------|---------------------|--------------------------|
| PA Chest | 30 | 15 | 23 |
| AP Cervical Spine | 250 | 175 | 13 |
| AP Thoracic Spine | 900 | 600 | 23 |
| AP Lumbar Spine | 1000 | 675 | 23 |
| AP Abdomen | 750 | 500 | 23 |
| AP Retrograde Pyelogram | 900 | 600 | 23 |
| Lateral Skull | 300 | 200 | 15 |
| Dental (bitewing or periapical) | 700 | 350 | not applicable |

OR

| Examination | Dose mrem maximum | Dose mrem recommended | Body part thickness (cm) |
|---------------------------------|-------------------|-----------------------|--------------------------|
| PA Chest | 3.13 | 1.57 | 23 |
| AP Cervical Spine | 10.88 | 7.61 | 13 |
| AP Thoracic Spine | 93.96 | 62.64 | 23 |
| AP Lumbar Spine | 104.4 | 70.47 | 23 |
| AP Abdomen | 78.3 | 52.2 | 23 |
| AP Retrograde Pyelogram | 93.96 | 62.64 | 23 |
| Lateral Skull | 6.54 | 4.36 | 15 |
| Dental (bitewing or periapical) | 1.05 | 0.53 | not applicable |

| | |
|---|---|
| | <p>ESE for all x-ray units in facility are not within 20 percent of one another.</p> <p>Typical exposure value for the x-ray unit is not posted</p> <p>Exposure reproducibility is greater than 5%</p> |
| Public exposure | <p>Public exposure exceeded - 100 millirem per year</p> <p>Public is not protected from scatter radiation</p> |
| Operator conditions | <p>Operator exposure exceeded - 5000 millirem per year</p> <p>Operator cannot observe patient during exposure</p> <p>Operator cannot monitor kVp, mA, time, mAs during exposure</p> <p>Operator is not protected during exposure</p> <p>Satisfactory lead gloves are not available</p> <p>Mobile or stationary exposure switch cord is less than 6 feet long</p> <p>Exposure switch not located to prevent x-ray activation when operator is outside of of the control booth</p> <p>Operator holds film in patient's mouth</p> |
| Physical condition (x-ray unit, shielding, etc.) | <p>Console does not indicate tubes for multiple setup</p> <p>Panoramic or 3D unit does not reset before restarting</p> <p>Motion of panoramic or 3D unit is not smooth or is impeded</p> <p>X-ray tube head locks into position for panoramic, cephalometric and or 3D unit</p> <p>Table locks, tube crane locks, bucky-cassette locks are not functioning</p> <p>Filters for soft tissue imaging for cephalometric imaging are not available</p> <p>Focal spot is not indicated on the x-ray tube</p> <p>Source to image distance is less than 7 7/8 inches for intra-oral x-ray tubes</p> <p>Source to image distance is less than 40 inches for medical x-ray machines</p> <p>Unit is inaccurate/not calibrated in terms of examination distance</p> <p>Tube head is unstable</p> <p>Overhead crane does not move easily</p> <p>Exposure switch is not labeled</p> <p>Unit does not have visual indication of kVp, mA, time, or mAs</p> <p>Unit does not have audible/visual indication of exposure</p> <p>Angulation indicator on x-ray unit is not functioning</p> <p>Typical exposure for x-ray unit is not posted</p> <p>Structural shielding is inadequate</p> <p>Door interlock system is not functioning</p> <p>Condition of high voltage and other cables is inadequate</p> <p>X-ray head leaks oil</p> <p>Wires are exposed on tube head</p> <p>X-ray exposure button is missing or broken</p> <p>Wires are exposed on exposure switch</p> <p>Preventive maintenance records for x-ray machines and processor are not kept</p> <p>Bare sheet lead on walls/doors is not covered</p> |


| |
|--|
| X-ray unit is not registered |
| Vermont State licenses are not displayed |
| No documentation of LMP (chiropractic) |
| Repeat rate analysis is not performed (chiropractic) |

SUMMARY OF ALL INSPECTIONS


Total Number of Inspections Performed 127
Total Number of Facilities not in Compliance 56

| | |
|--|-------|
| TOTAL NONCOMPLIANCES | 91 |
| Average noncompliances per noncompliant facility | 1.63 |
| Range of number of noncompliances/facility | 0 - 8 |

| TOTAL FACILITY NONCOMPLIANCES | 73 | PERCENTAGE OF TOTAL FACILITY NONCOMPLIANCES |
|--------------------------------------|----|--|
| 1 Film/Screen | 10 | 13.7 |
| 2 Processing | 6 | 8.2 |
| 3 Darkroom/Safelight | 11 | 15.1 |
| 4 Personnel Monitoring | 9 | 12.3 |
| 5 Patient Shielding | 35 | 47.9 |
| 6 License Not Displayed | 0 | 0.0 |
| 7 Repeat Analysis Not Performed | 2 | 2.7 |



| TOTAL RADIOGRAPHIC NONCOMPLIANCES | 18 | PERCENTAGE OF TOTAL RADIOGRAPHIC NONCOMPLIANCES |
|--|----|---|
| 1 Collimation | 0 | 0.0 |
| 2 Timer | 2 | 11.1 |
| 3 kVp & Filtration | 5 | 27.8 |
| 4 Patient entrance skin exposure | 0 | 0.0 |
| 5 Public exposure | 0 | 0.0 |
| 6 Operator conditions | 2 | 11.1 |
| 7 Physical condition (x-ray unit, shielding) | 9 | 50.0 |
| 8 Unit not registered | 0 | 0.0 |



| Annual Dose to Occupational Worker | | | |
|---|----------------------------------|--------------------------------|--------------------------------------|
| Type of Facility | Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
| Dental | 2.4 | 0.0002 - 64 | 5000 |
| Medical | 0.25 | 0.0029 - 81 | 5000 |
| Chiropractic | 0.027 | 0.004 - 0.063 | 5000 |
| Podiatric | 0.028 | na | 5000 |
| Veterinary | 0.37 | 0.0001 - 6.3 | 5000 |

| Annual Dose to Public | | | |
|------------------------------|----------------------------------|--------------------------------|--------------------------------------|
| Type of Facility | Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
| Dental | 7.3 | .0043 - 104* | 100 |
| Medical | 2.9 | 0.026 - 53 | 100 |
| Chiropractic | 0.43 | 0.0002 - 2.1 | 100 |
| Podiatric | 0.13 | 0.0083 - 0.25 | 100 |
| Veterinary | 1.1 | 0.0012 - 33 | 100 |

*The dose to the public at this facility was measured at a distance of less than usual 6 feet due to the layout of the facility. The facility does not allow members of the public in that area during an x-ray examination.

DENTAL INSPECTIONS


Total Number of Inspections Performed 78
Total Number of Facilities not in Compliance 34
Non-compliance Items

| | |
|--|------|
| TOTAL NONCOMPLIANCES | 53 |
| Average noncompliances per noncompliant facility | 1.56 |
| Range of number of noncompliances | 0-8 |


| TOTAL FACILITY NONCOMPLIANCES | 39 | PERCENTAGE OF TOTAL FACILITY NONCOMPLIANCES |
|--------------------------------------|----|--|
| Film/Screen | 5 | 12.8 |
| Processing | 4 | 10.3 |
| Darkroom/Safelight | 8 | 20.5 |
| Personnel Monitoring | 0 | 0.0 |
| Patient Shielding | 22 | 56.4 |

| TOTAL RADIOGRAPHIC NONCOMPLIANCES | 14 | PERCENTAGE OF TOTAL RADIOGRAPHIC NONCOMPLIANCES |
|--|----|--|
| Collimation | 0 | 0.0 |
| Timer | 2 | 14.3 |
| kVp & Filtration | 3 | 21.4 |
| Patient entrance skin exposure | 0 | 0.0 |
| Public exposure | 0 | 0.0 |
| Operator conditions | 1 | 7.1 |
| Physical condition (x-ray unit, shielding) | 8 | 57.1 |
| Unit not registered | 0 | 0.0 |

Percent of Total Dental Facility Noncompliances



Percent of Total Dental Radiographic Noncompliances



Dose to Patients Per Exposure

| Exam Type | Average millirem per exposure | Range millirem per exposure | Vermont State maximum dose millirem | Vermont State recommended dose millirem | NCRP DRL* millirem |
|-----------------------------|--------------------------------------|------------------------------------|--|--|---------------------------|
| Intra-oral D speed film | 0.47 | 0.27 - 0.69 | 1.05 | 0.53 | 0.28 |
| Intra-oral E speed film | 0.3 | 0.16 - 0.45 | 1.05 | 0.53 | 0.28 |
| Intra-oral F speed film | 0.26 | 0.10 - 0.58 | 1.05 | 0.53 | 0.28 |
| Intra-oral Portable digital | 0.07 | 0.04 - 0.11 | 1.05 | 0.53 | 0.28 |
| Intra-oral CR digital | 0.25 | 0.05 - 0.50 | 1.05 | 0.53 | 0.28 |
| Intra-oral DR digital | 0.15 | 0.06 - 0.54 | 1.05 | 0.53 | 0.28 |
| Panoramic film | 1.05 | 0.48 - 2.16 | na** | na | na |
| Panoramic digital | 0.86 | 0.14 - 1.62 | na | na | na |
| Cephalometric | 0.02 | 0.015 - 0.032 | na | na | 0.024 |
| Cephalometric digital | na | na | na | na | 0.024 |
| Cephalometric scanner | 0.33 | 0.20 - 0.46 | na | na | 0.024 |
| 3 Dimensional | 0.54 | 0.29 - 1.35 | na | na | na |

* DRL = Diagnostic Reference Level (derived from NEXT data) adjusted to millirem, NCRP Report 145, 2003

** na = not applicable

Annual Dose to Occupational Worker

| Exam Type | Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|-----------------------------|----------------------------------|--------------------------------|--------------------------------------|
| Intra-oral D speed film | 1.6 | 0.024 - 3.5 | 5000 |
| Intra-oral E speed film | 0.81 | 0.004 - 2.4 | 5000 |
| Intra-oral F speed film | 4.1 | 0.004 - 20 | 5000 |
| Intra-oral Portable digital | 0.27 | 0.10 - 0.43 | 5000 |
| Intra-oral CR digital | 1.6 | 0.011 - 7.9 | 5000 |
| Intra-oral DR digital | 2.1 | 0.001 - 41 | 5000 |
| Panoramic film | 0.6 | 0.007 - 4.1 | 5000 |
| Panoramic digital | 3.1 | 0.007 - 25 | 5000 |
| Cephalometric | 0.013 | 0.001 - 0.025 | 5000 |
| Cephalometric digital | na | na | 5000 |
| Cephalometric scanner | 2.5 | 0.32 - 4.1 | 5000 |
| 3 Dimensional | 2.9 | 0.12 - 11 | 5000 |

Annual Dose to Public

| Exam Type | Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|--------------------------------|----------------------------------|--------------------------------|--------------------------------------|
| Intra-oral D speed film | 16.3 | 0.61 - 61 | 100 |
| Intra-oral E speed film | 7.9 | 0.012 - 6.7 | 100 |
| Intra-oral F speed film | 11 | 0.10 - 55 | 100 |
| Intra-oral Portable CR digital | na | na | 100 |
| Intra-oral CR digital | 5 | 0.24 - 14 | 100 |
| Intra-oral DR digital | 6 | 0.035 - 42 | 100 |
| Panoramic film | 2.5 | 0.022 - 13 | 100 |
| Panoramic digital | 10 | 0.022 - 81 | 100 |
| Cephalometric | 0.06 | 0.004 - 0.11 | 100 |
| Cephalometric digital | na | na | 100 |
| Cephalometric scanner | 5.4 | 1.5 - 13 | 100 |
| 3 Dimensional | 3.5 | 0.14 - 104* | 100 |

*The dose to the public at this facility was measured at a distance of less than usual 6 feet due to the layout of the facility. The facility does not allow members of the public in that area during an x-ray examination.

MEDICAL INSPECTIONS


Total Number of Inspections Performed 12
Total Number of Facilities not in Compliance 4
Non-compliance Items

| | |
|--|------|
| TOTAL NONCOMPLIANCES | 5 |
| Average noncompliances per noncompliant facility | 1.25 |
| Range of number of noncompliances | 0-2 |


| TOTAL FACILITY NONCOMPLIANCES | 3 | PERCENTAGE OF TOTAL FACILITY NONCOMPLIANCES |
|--------------------------------------|---|--|
| Film/Screen | 0 | 0 |
| Processing | 0 | 0 |
| Darkroom/Safelight | 1 | 33.3 |
| Personnel Monitoring | 1 | 33.3 |
| Patient Shielding | 1 | 33.3 |

| TOTAL RADIOGRAPHIC NONCOMPLIANCES | 2 | PERCENTAGE OF TOTAL RADIOGRAPHIC NONCOMPLIANCES |
|--|---|--|
| Collimation | 0 | 0 |
| Timer | 0 | 0 |
| kVp & Filtration | 1 | 50 |
| Patient entrance skin exposure | 0 | 0 |
| Public exposure | 0 | 0 |
| Operator conditions | 0 | 0 |
| Physical condition (x-ray unit, shielding) | 1 | 50 |
| Unit not registered | 0 | 0 |

Percent of Total Medical Facility Noncompliances



Percent of Total Medical Radiographic Noncompliances



Dose to Patients Per Exposure

| Type of Exam | Average millirem per exposure | Range millirem per exposure | Vermont State maximum dose millirem | Vermont State recommended dose millirem | NCRP *DRL millirem |
|-----------------------|-------------------------------|-----------------------------|-------------------------------------|---|--------------------|
| PA Chest | 1.8 | 0.63 - 2.7 | 3.13 | 1.57 | 1.8 |
| AP Cervical Spine | 4 | 0.96 - 7.1 | 10.88 | 7.61 | na |
| AP Thoracic Spine | 13 | na | 93.96 | 62.64 | na |
| AP Lumbar Spine | 31 | 20 - 98 | 104.4 | 70.47 | 50 |
| AP Abdomen | na | na | 78.3 | 52.2 | 41 |
| AP Retrograde | na | na | 93.96 | 62.64 | na |
| Lateral Skull | 2.8 | na | 6.54 | 4.36 | na |
| Hand | 0.13 | 0.05 - 0.25 | na | na | na |
| Wrist | 0.05 | 0.03 - 0.08 | na | na | na |
| Arm | na | na | na | na | na |
| Shoulder | 5 | 3.3 - 6.7 | na | na | na |
| Leg | na | na | na | na | na |
| Knee | 4.1 | 1.3 - 4.7 | na | na | na |
| Ankle | 0.21 | 0.17 - 0.26 | na | na | na |
| DP Foot | 0.12 | na | na | na | na |
| Lateral Foot | na | na | na | na | na |
| Fluoroscopy | | | na | na | na |
| Arm | na | na | na | na | na |
| Knee | na | na | na | na | na |
| Ankle | na | na | na | na | na |
| AP Cervical | na | na | na | na | na |
| AP Lumbar | 98 | na | na | na | na |
| Fluoroscopy Spot Film | 1.3 | na | na | na | na |
| Sinus | na | na | na | na | na |

* DRL = Diagnostic Reference Level (derived from NEXT data) adjusted to millirem, NCRP Report 172, 2012

Annual Dose to Occupational Worker

| Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|--|--|--|
| 0.25 | 0.003 - 81 | 5000 |

Annual Dose to Public

| Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|--|--|--|
| 2.9 | 0.026 - 53 | 100 |

CHIROPRACTIC INSPECTIONS


Total Number of Inspections Performed 8
Total Number of Facilities not in Compliance 3
Non-compliance Items

| | |
|--|-------|
| TOTAL NONCOMPLIANCES | 10 |
| Average noncompliances per noncompliant facility | 3.33 |
| Range of number of noncompliances | 0 - 5 |


| TOTAL FACILITY NONCOMPLIANCES | 10 | PERCENTAGE OF TOTAL FACILITY NONCOMPLIANCES |
|--------------------------------------|----|--|
| Film/Screen | 1 | 10.0 |
| Processing | 1 | 10.0 |
| Darkroom/Safelight | 1 | 10.0 |
| Personnel Monitoring | 0 | 0.0 |
| Patient Shielding | 5 | 50.0 |
| License Displayed | 0 | 0.0 |
| Repeat Analysis | 2 | 20.0 |

| TOTAL RADIOGRAPHIC NONCOMPLIANCES | 0 | PERCENTAGE OF TOTAL RADIOGRAPHIC NONCOMPLIANCES |
|--|---|--|
| Collimation | 0 | 0.0 |
| Timer | 0 | 0.0 |
| kVp & Filtration | 0 | 0.0 |
| Patient entrance skin exposure | 0 | 0.0 |
| Public exposure | 0 | 0.0 |
| Operator conditions | 1 | 0.0 |
| Physical condition (x-ray unit, shielding) | 0 | 0.0 |
| Unit not registered | 0 | 0.0 |

Percent of Total Chiropractic Facility Noncompliances



Percent of Total Chiropractic Radiographic Noncompliances



Dose to Patients Per Exposure

| Type of Exam | Average millirem per exposure | Range millirem per exposure | Vermont State maximum dose millirem | Vermont State recommended dose millirem | NCRP *DRL millirem |
|-------------------|-------------------------------|-----------------------------|-------------------------------------|---|--------------------|
| PA Chest | na | na | 3.13 | 1.57 | 1.8 |
| AP Cervical Spine | 2.5 | 1.7 - 3.7 | 10.88 | 7.61 | na |
| AP Thoracic Spine | 26 | 13 - 46 | 93.96 | 62.64 | na |
| AP Lumbar Spine | 40 | 8.9 - 70 | 104.4 | 70.47 | 50 |
| AP Abdomen | na | na | 78.3 | 52.2 | 41 |
| AP Retrograde | na | na | 93.96 | 62.64 | na |
| Lateral Skull | na | na | 6.54 | 4.36 | na |

| Type of Exam | Average millirem per exposure | Range millirem per exposure | Vermont State maximum dose millirem | Vermont State recommended dose millirem | NCRP *DRL millirem |
|--------------|-------------------------------|-----------------------------|-------------------------------------|---|--------------------|
| Hand | na | na | na | na | na |
| Wrist | na | na | na | na | na |
| Arm | na | na | na | na | na |
| Shoulder | na | na | na | na | na |
| Leg | na | na | na | na | na |
| Knee | 0.43 | na | na | na | na |
| Ankle | 0.026 | na | na | na | na |
| DP Foot | na | na | na | na | na |
| Lateral Foot | na | na | na | na | na |

* DRL = Diagnostic Reference Level (derived from NEXT data) adjusted to millirem, NCRP Report 172, 2012

Annual Dose to Occupational Worker

| Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|---------------------------|-------------------------|-------------------------------|
| 0.027 | 0.004 - 0.063 | 5000 |

Annual Dose to Public

| Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|---------------------------|-------------------------|-------------------------------|
| 0.43 | 0.0002 - 2.1 | 100 |

PODIATRIC INSPECTIONS

Total Number of Inspections Performed 2
Total Number of Facilities not in Compliance 0
Non-compliance Items

| | |
|--|---|
| TOTAL NONCOMPLIANCES | 0 |
| Average number noncompliances per facility | 0 |
| Range of number of noncompliances | 0 |


| TOTAL FACILITY NONCOMPLIANCES | 0 | PERCENTAGE OF TOTAL FACILITY NONCOMPLIANCES |
|--------------------------------------|---|--|
| Film/Screen | 0 | 0.0 |
| Processing | 0 | 0.0 |
| Darkroom/Safelight | 0 | 0.0 |
| Personnel Monitoring | 0 | 0.0 |
| Patient Shielding | 0 | 0.0 |

| TOTAL RADIOGRAPHIC NONCOMPLIANCES | 0 | PERCENTAGE OF TOTAL RADIOGRAPHIC NONCOMPLIANCES |
|--|---|--|
| Collimation | 0 | 0.0 |
| Timer | 0 | 0.0 |
| kVp & Filtration | 0 | 0.0 |
| Patient entrance skin exposure | 0 | 0.0 |
| Public exposure | 0 | 0.0 |
| Operator conditions | 0 | 0.0 |
| Physical condition (x-ray unit, shielding) | 0 | 0.0 |
| Unit not registered | 0 | 0.0 |

Percent of Total Podiatric Facility Noncompliances



Percent of Total Podiatric Radiographic Noncompliances



Dose to Patients Per Exposure

| Type of Exam | Average millirem per exposure | Range millirem per exposure | Vermont State maximum dose millirem | Vermont State recommended dose millirem | NCRP *DRL millirem |
|--------------|-------------------------------|-----------------------------|-------------------------------------|---|--------------------|
| DP Foot | 0.14 | na | na | na | na |
| Lateral Foot | 0.24 | 0.16 - 0.31 | na | na | na |

Annual Dose to Occupational Worker

| Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|---------------------------|-------------------------|-------------------------------|
| 0.028 | na | 5000 |

Annual Dose to Public

| Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|---------------------------|-------------------------|-------------------------------|
| 0.13 | 0.008 - 0.25 | 100 |

VETERINARIAN INSPECTIONS

Total Number of Inspections Performed 26

Total Number of Facilities not in Compliance 15


Non-compliance Items

| | |
|--|-------|
| TOTAL NONCOMPLIANCES | 23 |
| Average number noncompliances per facility | 1.53 |
| Range of number of noncompliances | 0 - 3 |


| TOTAL FACILITY NONCOMPLIANCES | 21 | PERCENTAGE OF TOTAL FACILITY NONCOMPLIANCES |
|--------------------------------------|----|--|
| Film/Screen | 4 | 19.0 |
| Processing | 1 | 4.8 |
| Darkroom/Safelight | 1 | 4.8 |
| Personnel Monitoring | 8 | 38.1 |
| Patient Shielding | 7 | 33.3 |

| TOTAL RADIOGRAPHIC NONCOMPLIANCES | 2 | PERCENTAGE OF TOTAL RADIOGRAPHIC NONCOMPLIANCES |
|--|---|--|
| Collimation | 0 | 0 |
| Timer | 0 | 0 |
| kVp & Filtration | 1 | 50 |
| Patient entrance skin exposure | 0 | 0 |
| Public exposure | 0 | 0 |
| Operator conditions | 1 | 50 |
| Physical condition (x-ray unit, shielding) | 0 | 0 |
| Unit not registered | 0 | 0 |

Percent of Total Veterinary Facility Noncompliances



Percent of Total Veterinary Radiographic Noncompliances



Exposure to Animals Per Exam

| Type of Exam | Average milliroentgen per exposure | Range milliroentgen per exposure |
|-----------------------------|---|---|
| Dog chest | 31 | 11 - 145 |
| Dog abdomen | 46 | 10 - 124 |
| Dog extremity | 7.3 | 1.2 - 23 |
| Dog dental | 127 | 91 - 192 |
| Cat o-gram | 11 | 2.8 - 48 |
| Cat chest/abdomen | 8 | 7.7 - 8.6 |
| Cat extremity | 7.4 | 2.4 - 14 |
| Cat dental | 71 | 50 - 89 |
| Horse hoof | 23 | 6.9 - 153 |
| Horse navicular | 25 | 9.6 - 153 |
| Horse fetlock/pastern/ankle | 23 | 9.6 - 153 |
| Horse carpus/knee | 25 | 9.6 - 153 |
| Horse hock | 27 | 9.6 - 153 |
| Horse gaskin/forearm | 26 | 9.6 - 153 |
| Horse canon | 27 | 9.6 - 153 |
| Horse stifle/hip | 49 | 21 - 153 |
| Horse spine | 28 | 21 - 153 |

Annual Dose to Occupational Worker

| STATIONARY X-RAY Position of Operator | Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|--|--|--|--|
| Operator exposure at edge of table | 11 | 0.21 - 45 | 5000 |
| Operator exposure at opposite ends of table | 6.2 | 0.55 - 28 | 5000 |
| Operator exposure 3 feet from x-ray unit | 3.3 | 0.10 - 11 | 5000 |
| Operator exposure 6 feet from x-ray unit | 0.93 | 0.025 - 3.8 | 5000 |
| Operator exposure behind shield, wall, or door | 0.42 | 0.003 - 6.3 | 5000 |
| Extremity exposure | 36 | 0.54 - 163 | 50,000 |

| PORTABLE X-RAY Position of Operator | Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|--|--|--|--|
| Operator exposure holding unit | 7.7 | 0.029 - 43 | 5000 |
| Operator exposure 3 feet from x-ray unit | 3.9 | 0.027 - 41 | 5000 |
| Operator exposure 6 feet from x-ray unit | 0.39 | 0.01 - 2.1 | 5000 |
| Operator exposure 9 feet from x-ray unit | na | na | 5000 |
| Operator exposure at end of exposure cord | na | na | 5000 |
| Operator exposure behind shield, wall, or door | na | na | 5000 |
| Extremity exposure | 9.1 | 0.4 - 55 | 50,000 |

| DENTAL X-RAY Position of Operator | Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|--|--|--|--|
| Operator exposure at edge of table | 0.76 | 0.019 - 1.4 | 5000 |
| Operator exposure 6 feet from x-ray unit | 0.06 | 0.021 - 0.11 | 5000 |
| Operator exposure at end of exposure cord | na | na | 5000 |
| Operator exposure behind shield, wall, or door | 0.008 | 0.0001 - 0.021 | 5000 |
| Extremity exposure | 2.9 | 0.033 - 5.8 | 50,000 |

Annual Dose to Public

| | Average millirem per year | Range millirem per year | Maximum Allowable millirem/yr |
|------------------|--|--|--|
| Stationary X-Ray | 0.22 | 0.0012 - 1.3 | 100 |
| Portable X-Ray | 2.1 | 0.0017 - 0.42 | 100 |
| Dental X-Ray | 6.6 | 0.0003 - 33 | 100 |